

Porlock Parish

Clerk to the Council
Mrs. D Taylor-Pitkin
Office Opening Hours:

Monday	CLOSED
Tuesday	9:30 – 14:00
Wednesday	9:30 – 14:00
Thursday	9:30 – 14:00
Friday	CLOSED



Council

Contact Details:
Parish Office
Old School Centre, West End
Porlock, Somerset.
TA24 8QD
Office Tel: 01643 863350
Email: porlockpc@gmail.com
<https://porlockparishcouncil.org>

Porlock Abbeyfield Defibrillator.

Monthly checks have been conducted on 1-2nd of each month January - May 2021. No interference has been identified. The equipment is complete and in order. The battery and pad pack was replaced in January 2021 (£125.75). These will expire again in December 2023. The equipment was unserviceable between 29th March and 6th April during building works at Abbeyfield. South West Ambulance Authority have been updated throughout and supplied with the results of monthly service checks.

Porlock Weir Car Park Defibrillator.

Monthly checks have been conducted on 1-2nd of each month January - May 2021. No interference has been identified. The equipment is complete and in order. The battery and pad pack was replaced in March 2021 (£125.75). These will expire again in March 2024. South West Ambulance Authority have been updated throughout and supplied with the results of monthly service checks.

General

As you are aware a written defibrillator maintenance procedure has been compiled and circulated. Nigel Lamacraft is also familiar with how to conduct the monthly maintenance checks.

Since it is more than two years since the defibrillators were introduced to Porlock community, it might be considered a suitable time to determine whether and how residents could, in anticipation of an emergency, be reminded of the key elements of their use.

